

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040760.D
 Acq On : 21 Mar 2024 21:58
 Operator : JC/MD
 Sample : PB159677ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159677ZHE#11

Quant Time: Mar 22 03:02:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	41808	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	74262	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	69643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	36833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35066	49.882	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.760%
35) Dibromofluoromethane	5.391	113	25219	45.385	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.760%
50) Toluene-d8	8.647	98	90329	47.475	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.960%
62) 4-Bromofluorobenzene	11.079	95	34979	46.532	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.060%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	27203	109.249	ug/l	98
20) Methylene Chloride	2.794	84	2340	4.299	ug/l	95
25) 2-Butanone	4.574	43	2990	7.667	ug/l #	89
43) Isopropyl Acetate	6.336	43	39371	26.987	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
Data File : VX040760.D
Acq On : 21 Mar 2024 21:58
Operator : JC/MD
Sample : PB159677ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159677ZHE#11

Quant Time: Mar 22 03:02:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:41:46 2024
Response via : Initial Calibration

