

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040762.D
 Acq On : 21 Mar 2024 22:43
 Operator : JC/MD
 Sample : PB159677ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159677ZHE#13

Quant Time: Mar 22 03:03:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	41125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	73496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	70016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	35301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	34099	49.312	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.620%
35) Dibromofluoromethane	5.391	113	24815	45.123	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.240%
50) Toluene-d8	8.646	98	89062	47.297	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.600%
62) 4-Bromofluorobenzene	11.079	95	34693	46.633	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.260%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	28571	116.649	ug/l	93
20) Methylene Chloride	2.794	84	2136	3.990	ug/l	96
25) 2-Butanone	4.562	43	2653	6.915	ug/l	91
43) Isopropyl Acetate	6.336	43	39463	27.332	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
Data File : VX040762.D
Acq On : 21 Mar 2024 22:43
Operator : JC/MD
Sample : PB159677ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159677ZHE#13

Quant Time: Mar 22 03:03:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:41:46 2024
Response via : Initial Calibration

