

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040768.D
 Acq On : 22 Mar 2024 01:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 03:06:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.647	0.645	0.3	84	0.00
3 P	Chloromethane	0.662	0.565	14.7	77	0.00
4 C	Vinyl Chloride	0.627	0.596	4.9#	80	0.00
5 T	Bromomethane	0.422	0.404	4.3	71	0.00
6 T	Chloroethane	0.440	0.366	16.8	72	0.00
7 T	Trichlorofluoromethane	1.096	1.020	6.9	79	0.00
8 T	Diethyl Ether	0.352	0.342	2.8	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.631	0.566	10.3	77	0.00
10 T	Methyl Iodide	0.835	0.779	6.7	82	0.00
11 T	Tert butyl alcohol	0.135	0.148	-9.6	98	0.00
12 CM	1,1-Dichloroethene	0.591	0.535	9.5#	79	0.00
13 T	Acrolein	0.080	0.076	5.0	89	0.00
14 T	Allyl chloride	1.058	0.984	7.0	81	0.00
15 T	Acrylonitrile	0.332	0.339	-2.1	87	0.00
16 T	Acetone	0.298	0.295	1.0	89	0.00
17 T	Carbon Disulfide	1.696	1.435	15.4	75	0.00
18 T	Methyl Acetate	0.798	0.762	4.5	89	0.00
19 T	Methyl tert-butyl Ether	2.038	2.009	1.4	85	0.00
20 T	Methylene Chloride	0.651	0.610	6.3	82	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.573	8.0	81	0.00
22 T	Diisopropyl ether	2.065	2.036	1.4	84	0.00
23 T	Vinyl Acetate	1.952	1.939	0.7	84	0.00
24 P	1,1-Dichloroethane	1.186	1.138	4.0	83	0.00
25 T	2-Butanone	0.466	0.490	-5.2	90	0.00
26 T	2,2-Dichloropropane	1.035	0.863	16.6	70	0.00
27 T	cis-1,2-Dichloroethene	0.722	0.690	4.4	84	0.00
28 T	Bromochloromethane	0.543	0.560	-3.1	99	0.00
29 T	Tetrahydrofuran	0.310	0.324	-4.5	92	0.00
30 C	Chloroform	1.270	1.239	2.4#	84	0.00
31 T	Cyclohexane	0.999	0.925	7.4	79	0.00
32 T	1,1,1-Trichloroethane	1.092	1.098	-0.5	83	0.01
33 S	1,2-Dichloroethane-d4	0.841	0.886	-5.4	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.374	0.358	4.3	95	0.00
36 T	1,1-Dichloropropene	0.551	0.478	13.2	84	0.00
37 T	Ethyl Acetate	0.626	0.578	7.7	86	0.01
38 T	Carbon Tetrachloride	0.620	0.541	12.7	84	0.00
39 T	Methylcyclohexane	0.610	0.567	7.0	84	0.00
40 TM	Benzene	1.521	1.384	9.0	92	0.00
41 T	Methacrylonitrile	0.323	0.323	0.0	90	0.00
42 TM	1,2-Dichloroethane	0.616	0.605	1.8	98	0.00
43 T	Isopropyl Acetate	0.982	0.964	1.8	92	0.00
44 TM	Trichloroethene	0.365	0.346	5.2	89	0.00
45 C	1,2-Dichloropropane	0.357	0.344	3.6#	89	0.00
46 T	Dibromomethane	0.283	0.283	0.0	91	0.00
47 T	Bromodichloromethane	0.578	0.548	5.2	86	0.00
48 T	Methyl methacrylate	0.449	0.458	-2.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040768.D
 Acq On : 22 Mar 2024 01:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 03:06:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	105	0.00
50 S	Toluene-d8	1.281	1.284	-0.2	100	0.00
51 T	4-Methyl-2-Pentanone	0.609	0.622	-2.1	94	0.00
52 CM	Toluene	0.907	0.854	5.8#	85	0.00
53 T	t-1,3-Dichloropropene	0.597	0.575	3.7	89	0.00
54 T	cis-1,3-Dichloropropene	0.618	0.595	3.7	87	0.00
55 T	1,1,2-Trichloroethane	0.369	0.346	6.2	88	0.00
56 T	Ethyl methacrylate	0.618	0.608	1.6	87	0.00
57 T	1,3-Dichloropropane	0.630	0.572	9.2	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.324	0.302	6.8	96	0.00
59 T	2-Hexanone	0.472	0.467	1.1	93	0.00
60 T	Dibromochloromethane	0.434	0.420	3.2	89	0.00
61 T	1,2-Dibromoethane	0.390	0.366	6.2	87	0.00
62 S	4-Bromofluorobenzene	0.506	0.509	-0.6	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.372	0.307	17.5	81	0.00
65 PM	Chlorobenzene	1.134	1.040	8.3	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.358	7.3	84	0.00
67 C	Ethyl Benzene	2.031	1.872	7.8#	85	0.00
68 T	m/p-Xylenes	0.744	0.690	7.3	86	0.00
69 T	o-Xylene	0.714	0.660	7.6	84	0.00
70 T	Styrene	1.196	1.146	4.2	87	0.00
71 P	Bromoform	0.315	0.286	9.2	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.794	3.320	12.5	87	0.00
74 T	N-amyl acetate	1.868	1.758	5.9	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.284	1.164	9.3	91	0.00
76 T	1,2,3-Trichloropropane	1.075	1.032	4.0	92	0.00
77 T	Bromobenzene	0.918	0.819	10.8	89	0.00
78 T	n-propylbenzene	4.678	4.303	8.0	91	0.00
79 T	2-Chlorotoluene	2.699	2.590	4.0	95	0.00
80 T	1,3,5-Trimethylbenzene	3.180	3.040	4.4	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.425	0.369	13.2	84	0.00
82 T	4-Chlorotoluene	3.385	3.077	9.1	92	0.00
83 T	tert-Butylbenzene	3.190	2.931	8.1	88	0.00
84 T	1,2,4-Trimethylbenzene	3.201	3.010	6.0	92	0.00
85 T	sec-Butylbenzene	4.056	3.750	7.5	89	0.00
86 T	p-Isopropyltoluene	3.364	3.181	5.4	92	0.00
87 T	1,3-Dichlorobenzene	1.754	1.612	8.1	89	0.00
88 T	1,4-Dichlorobenzene	1.862	1.648	11.5	91	0.00
89 T	n-Butylbenzene	3.392	2.714	20.0	78	0.00
90 T	Hexachloroethane	0.586	0.488	16.7	75	0.00
91 T	1,2-Dichlorobenzene	1.723	1.426	17.2	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.321	0.297	7.5	92	0.00
93 T	1,2,4-Trichlorobenzene	1.224	1.069	12.7	84	0.00
94 T	Hexachlorobutadiene	0.483	0.383	20.7	77	0.00
95 T	Naphthalene	3.874	3.650	5.8	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
Data File : VX040768.D
Acq On : 22 Mar 2024 01:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 22 03:06:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:41:46 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.203	1.056	12.2	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6