

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
 Data File : VX045384.D
 Acq On : 21 Mar 2025 13:41
 Operator : JC/MD
 Sample : Q1582-10DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20250313DL

Quant Time: Mar 21 23:14:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63633	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56632	55.951	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.900%
35) Dibromofluoromethane	5.385	113	43114	52.438	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.880%
50) Toluene-d8	8.647	98	150492	50.488	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.980%
62) 4-Bromofluorobenzene	11.079	95	50805	51.437	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.880%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	1899	2.568	ug/l	95
44) Trichloroethene	7.123	130	30716	36.740	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045384.D
Acq On : 21 Mar 2025 13:41
Operator : JC/MD
Sample : Q1582-10DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20250313DL

Quant Time: Mar 21 23:14:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

