

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
 Data File : VX045388.D
 Acq On : 21 Mar 2025 15:14
 Operator : JC/MD
 Sample : Q1582-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT119S1-20250313

Quant Time: Mar 21 23:16:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129665	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58910	55.281	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.560%
35) Dibromofluoromethane	5.385	113	45756	52.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.540%
50) Toluene-d8	8.647	98	159091	50.613	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.220%
62) 4-Bromofluorobenzene	11.079	95	49513	47.536	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.080%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	923	1.867	ug/l	96
44) Trichloroethene	7.141	130	1740	1.974	ug/l	88
64) Tetrachloroethene	9.268	164	393	0.534	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045388.D
Acq On : 21 Mar 2025 15:14
Operator : JC/MD
Sample : Q1582-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT119S1-20250313

Quant Time: Mar 21 23:16:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

