

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
 Data File : VX045389.D
 Acq On : 21 Mar 2025 15:38
 Operator : JC/MD
 Sample : Q1582-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20250314

Quant Time: Mar 21 23:16:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48608	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61331	56.901	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.800%
35) Dibromofluoromethane	5.385	113	46142	52.208	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.420%
50) Toluene-d8	8.647	98	159107	49.657	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	54884	51.692	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.380%
Target Compounds						
16) Acetone	2.386	43	1092	2.184	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045389.D
Acq On : 21 Mar 2025 15:38
Operator : JC/MD
Sample : Q1582-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20250314

Quant Time: Mar 21 23:16:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

