

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
 Data File : VX045391.D
 Acq On : 21 Mar 2025 16:25
 Operator : JC/MD
 Sample : Q1582-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT191D1-20250314

Quant Time: Mar 21 23:17:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133271	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62859	56.966	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.940%
35) Dibromofluoromethane	5.385	113	48629	54.569	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.140%
50) Toluene-d8	8.647	98	166764	51.618	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.240%
62) 4-Bromofluorobenzene	11.079	95	56902	53.151	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.300%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	1049	1.201	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.319	101	2443	3.030	ug/l	95
12) 1,1-Dichloroethene	2.313	96	769	0.902	ug/l #	83
16) Acetone	2.374	43	1135	2.217	ug/l #	84
44) Trichloroethene	7.123	130	11892	13.123	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045391.D
Acq On : 21 Mar 2025 16:25
Operator : JC/MD
Sample : Q1582-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT191D1-20250314

Quant Time: Mar 21 23:17:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

