

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
 Data File : VX045394.D
 Acq On : 21 Mar 2025 17:35
 Operator : JC/MD
 Sample : Q1608-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT172S1-20250317

Quant Time: Mar 21 23:18:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129916	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61439	58.340	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.680%
35) Dibromofluoromethane	5.379	113	46700	53.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.520%
50) Toluene-d8	8.647	98	164183	52.132	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.260%
62) 4-Bromofluorobenzene	11.079	95	58315	55.878	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.760%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	938	1.219	ug/l	96
16) Acetone	2.386	43	982	2.010	ug/l	99
44) Trichloroethene	7.117	130	29385	33.265	ug/l	96
52) Toluene	8.720	92	823	0.373	ug/l	86
64) Tetrachloroethene	9.269	164	1171	1.489	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045394.D
Acq On : 21 Mar 2025 17:35
Operator : JC/MD
Sample : Q1608-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT172S1-20250317

Quant Time: Mar 21 23:18:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

