

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
 Data File : VX045396.D
 Acq On : 21 Mar 2025 18:21
 Operator : JC/MD
 Sample : Q1608-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE139D2-20250317

Quant Time: Mar 21 23:19:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131424	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62856	59.503	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.000%	
35) Dibromofluoromethane	5.385	113	47872	54.474	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.940%	
50) Toluene-d8	8.647	98	163480	51.313	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.620%	
62) 4-Bromofluorobenzene	11.079	95	56775	53.778	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.560%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.318	101	5497	7.122	ug/l	98
12) 1,1-Dichloroethene	2.312	96	5185	6.354	ug/l	95
16) Acetone	2.379	43	886	1.808	ug/l	94
27) cis-1,2-Dichloroethene	4.501	96	1948	1.957	ug/l	95
38) Carbon Tetrachloride	5.678	117	2096	1.713	ug/l	90
44) Trichloroethene	7.116	130	180938	202.480	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045396.D
Acq On : 21 Mar 2025 18:21
Operator : JC/MD
Sample : Q1608-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE139D2-20250317

Quant Time: Mar 21 23:19:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

