



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032125\
Data File : VX045374.D
Acq On : 21 Mar 2025 09:29
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Title : SW846 8260
Last Update : Fri Feb 28 06:45:16 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	6834	PASS
75	95	30	60	55.8	16798	PASS
95	95	100	100	100.0	30088	PASS
96	95	5	9	6.4	1915	PASS
173	174	0.00	2	1.0	207	PASS
174	95	50	100	71.0	21352	PASS
175	174	5	9	7.5	1599	PASS
176	174	95	101	98.0	20923	PASS
177	176	5	9	6.3	1317	PASS