

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
 Data File : VX021358.D
 Acq On : 22 Mar 2021 15:15
 Operator : JC/MD
 Sample : M1735-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D1-20210316

Quant Time: Mar 23 02:01:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	79904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	130074	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113133	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	62474	47.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.34%
35) Dibromofluoromethane	5.464	113	44554	50.34	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.68%
50) Toluene-d8	8.694	98	158107	49.23	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.46%
62) 4-Bromofluorobenzene	11.118	95	55447	44.46	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.92%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.369	101	5153	6.17	ug/l	94
16) Acetone	2.428	43	1737	3.60	ug/l #	81
19) Methyl tert-butyl Ether	3.170	73	4712	1.54	ug/l	98
22) Diisopropyl ether	3.823	45	2594	0.80	ug/l #	84
24) 1,1-Dichloroethane	3.676	63	1405	0.81	ug/l #	89
27) cis-1,2-Dichloroethene	4.570	96	2505	2.40	ug/l	97
30) Chloroform	5.182	83	4201	2.32	ug/l	88
44) Trichloroethene	7.188	130	131277	136.84	ug/l	96
64) Tetrachloroethene	9.317	164	3148	3.73	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
Data File : VX021358.D
Acq On : 22 Mar 2021 15:15
Operator : JC/MD
Sample : M1735-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D1-20210316

Quant Time: Mar 23 02:01:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

