

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
 Data File : VX021363.D
 Acq On : 22 Mar 2021 17:05
 Operator : JC/MD
 Sample : M1735-18DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20210317DL

Quant Time: Mar 23 02:02:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	75176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	122893	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	107084	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	44834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	60172	48.80	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.60%
35) Dibromofluoromethane	5.464	113	42698	51.06	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.12%
50) Toluene-d8	8.694	98	152393	50.23	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.46%
62) 4-Bromofluorobenzene	11.117	95	54159	45.96	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.92%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.363	101	715	0.91	ug/l	93
27) cis-1,2-Dichloroethene	4.564	96	1109	1.13	ug/l	98
30) Chloroform	5.176	83	701	0.41	ug/l	89
44) Trichloroethene	7.188	130	49233	54.32	ug/l	95
64) Tetrachloroethene	9.317	164	2724	3.41	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
Data File : VX021363.D
Acq On : 22 Mar 2021 17:05
Operator : JC/MD
Sample : M1735-18DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20210317DL

Quant Time: Mar 23 02:02:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

