

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032221\  
 Data File : VX021365.D  
 Acq On : 22 Mar 2021 17:50  
 Operator : JC/MD  
 Sample : M1735-07DL 4X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE132D4-20210315DL

Quant Time: Mar 23 02:03:08 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031621W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Mar 16 14:00:21 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)     |
|-----------------------------|--------|-------|----------|----------|-------|--------------|
| -----                       |        |       |          |          |       |              |
| Internal Standards          |        |       |          |          |       |              |
| 1) Pentafluorobenzene       | 5.629  | 168   | 71779    | 50.00    | ug/l  | 0.00         |
| 34) 1,4-Difluorobenzene     | 6.829  | 114   | 118692   | 50.00    | ug/l  | 0.00         |
| 63) Chlorobenzene-d5        | 10.094 | 117   | 105274   | 50.00    | ug/l  | 0.00         |
| 72) 1,4-Dichlorobenzene-d4  | 12.059 | 152   | 44983    | 50.00    | ug/l  | 0.00         |
| System Monitoring Compounds |        |       |          |          |       |              |
| 33) 1,2-Dichloroethane-d4   | 6.029  | 65    | 57513    | 48.85    | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 97.70%       |
| 35) Dibromofluoromethane    | 5.464  | 113   | 41043    | 50.82    | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 101.64%      |
| 50) Toluene-d8              | 8.694  | 98    | 144592   | 49.34    | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 98.68%       |
| 62) 4-Bromofluorobenzene    | 11.118 | 95    | 51805    | 45.52    | ug/l  | 0.00         |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 91.04%       |
| Target Compounds            |        |       |          |          |       |              |
| 44) Trichloroethene         | 7.188  | 130   | 51155    | 58.43    | ug/l  | Qvalue<br>91 |
| -----                       |        |       |          |          |       |              |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032221\  
Data File : VX021365.D  
Acq On : 22 Mar 2021 17:50  
Operator : JC/MD  
Sample : M1735-07DL 4X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D4-20210315DL

Quant Time: Mar 23 02:03:08 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031621W.M  
Quant Title : SW846 8260  
QLast Update : Tue Mar 16 14:00:21 2021  
Response via : Initial Calibration

