

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
 Data File : VX021367.D
 Acq On : 22 Mar 2021 18:37
 Operator : JC/MD
 Sample : M1735-20DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20210317DL

Quant Time: Mar 23 02:03:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	71176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	116677	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	102653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	43054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	57027	48.85	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.70%
35) Dibromofluoromethane	5.464	113	41395	52.14	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.28%
50) Toluene-d8	8.694	98	143107	49.68	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.36%
62) 4-Bromofluorobenzene	11.118	95	50015	44.71	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.42%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.364	101	52296	70.30	ug/l	96
12) 1,1-Dichloroethene	2.358	96	370	0.51	ug/l #	73
44) Trichloroethene	7.188	130	4743	5.51	ug/l	91
64) Tetrachloroethene	9.318	164	1644	2.14	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
Data File : VX021367.D
Acq On : 22 Mar 2021 18:37
Operator : JC/MD
Sample : M1735-20DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20210317DL

Quant Time: Mar 23 02:03:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

