

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
 Data File : VX021368.D
 Acq On : 22 Mar 2021 19:01
 Operator : JC/MD
 Sample : M1735-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20210315DL

Quant Time: Mar 23 02:03:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	74860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	121538	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	105680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	42564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	59165	48.19	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.38%
35) Dibromofluoromethane	5.470	113	42326	51.18	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.36%
50) Toluene-d8	8.694	98	148537	49.50	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.00%
62) 4-Bromofluorobenzene	11.118	95	51049	43.81	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.62%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.364	101	39259	50.18	ug/l	95
12) 1,1-Dichloroethene	2.352	96	336	0.44	ug/l #	84
44) Trichloroethene	7.188	130	13472	15.03	ug/l	92
64) Tetrachloroethene	9.318	164	3650	4.62	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
Data File : VX021368.D
Acq On : 22 Mar 2021 19:01
Operator : JC/MD
Sample : M1735-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20210315DL

Quant Time: Mar 23 02:03:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

