

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
 Data File : VX021371.D
 Acq On : 22 Mar 2021 20:11
 Operator : JC/MD
 Sample : M1712-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RED-BRICK-COMP

Quant Time: Mar 23 02:04:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	72058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	119489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	107246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	44321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	58999	49.92	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.84%
35) Dibromofluoromethane	5.464	113	41636	51.21	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.42%
50) Toluene-d8	8.694	98	147300	49.93	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.86%
62) 4-Bromofluorobenzene	11.118	95	51842	45.25	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.50%
Target Compounds						
16) Acetone	2.423	43	9096	20.92	ug/l #	86
20) Methylene Chloride	2.835	84	4772	5.34	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032221\
Data File : VX021371.D
Acq On : 22 Mar 2021 20:11
Operator : JC/MD
Sample : M1712-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RED-BRICK-COMP

Quant Time: Mar 23 02:04:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

