

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027614.D
 Acq On : 22 Mar 2022 15:06
 Operator : JC/MD
 Sample : N2026-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 23 01:47:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	346015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	554020	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	483389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	208853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	191419	44.907	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.820%
35) Dibromofluoromethane	5.391	113	180027	50.506	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.020%
50) Toluene-d8	8.653	98	638321	47.324	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.640%
62) 4-Bromofluorobenzene	11.085	95	214939	43.355	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.700%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3047	1.754	ug/l	91
20) Methylene Chloride	2.788	84	1758	0.471	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027614.D
Acq On : 22 Mar 2022 15:06
Operator : JC/MD
Sample : N2026-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 23 01:47:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

