

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027617.D
 Acq On : 22 Mar 2022 16:17
 Operator : JC/MD
 Sample : N1998-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-141

Quant Time: Mar 23 01:47:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	331495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	541704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	498440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	221012	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	164297	40.233	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	80.460%
35) Dibromofluoromethane	5.391	113	179345	51.459	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.920%
50) Toluene-d8	8.653	98	641713	48.657	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.085	95	220038	45.392	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.780%
Target Compounds						
16) Acetone	2.386	43	3971	2.386	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027617.D
Acq On : 22 Mar 2022 16:17
Operator : JC/MD
Sample : N1998-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-141

Quant Time: Mar 23 01:47:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

