

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027618.D
 Acq On : 22 Mar 2022 16:40
 Operator : JC/MD
 Sample : N1998-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB031422

Quant Time: Mar 23 01:48:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	336336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	554023	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	496907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	230262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	193122	46.610	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.220%
35) Dibromofluoromethane	5.391	113	182121	51.093	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.180%
50) Toluene-d8	8.653	98	641188	47.536	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.080%
62) 4-Bromofluorobenzene	11.079	95	225437	45.472	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.940%
Target Compounds						
16) Acetone	2.386	43	5337	3.160	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027618.D
Acq On : 22 Mar 2022 16:40
Operator : JC/MD
Sample : N1998-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB031422

Quant Time: Mar 23 01:48:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

