

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027622.D
 Acq On : 22 Mar 2022 18:15
 Operator : JC/MD
 Sample : N1998-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Mar 23 01:49:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	342610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	554789	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	498067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	222042	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	192505	45.61	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.22%
35) Dibromofluoromethane	5.391	113	178707	50.07	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.14%
50) Toluene-d8	8.653	98	644686	47.73	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.46%
62) 4-Bromofluorobenzene	11.085	95	220494	44.41	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.82%
Target Compounds						
68) m/p-Xylenes	10.305	106	11604	1.68	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027622.D
Acq On : 22 Mar 2022 18:15
Operator : JC/MD
Sample : N1998-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Mar 23 01:49:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

