

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027623.D
 Acq On : 22 Mar 2022 18:39
 Operator : JC/MD
 Sample : N2028-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-109

Quant Time: Mar 23 01:49:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	335100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	549508	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	485252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	215590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	193132	46.785	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.560%
35) Dibromofluoromethane	5.391	113	176349	49.880	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.760%
50) Toluene-d8	8.653	98	632299	47.262	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.520%
62) 4-Bromofluorobenzene	11.079	95	218368	44.408	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.820%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	76216	96.876	ug/l #	72
16) Acetone	2.380	43	5175	3.076	ug/l	88
19) Methyl tert-butyl Ether	3.117	73	1800185	171.471	ug/l	100
22) Diisopropyl ether	3.770	45	3519	0.370	ug/l #	81
40) Benzene	6.037	78	7438	0.537	ug/l	98
67) Ethyl Benzene	10.195	91	6152	0.360	ug/l	98
68) m/p-Xylenes	10.305	106	4517	0.671	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	3774	0.286	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027623.D
Acq On : 22 Mar 2022 18:39
Operator : JC/MD
Sample : N2028-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-109

Quant Time: Mar 23 01:49:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

