

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027641.D
 Acq On : 23 Mar 2022 02:56
 Operator : JC/MD
 Sample : N2062-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Mar 23 07:42:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	347320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	592043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	535241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	234433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197567	46.175	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.360%
35) Dibromofluoromethane	5.391	113	184177	48.352	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.700%
50) Toluene-d8	8.653	98	637406	44.221	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.440%
62) 4-Bromofluorobenzene	11.085	95	237171	44.766	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.540%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4738	2.717	ug/l	91
19) Methyl tert-butyl Ether	3.123	73	28857	2.652	ug/l	100
85) sec-Butylbenzene	11.896	105	7378	0.429	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027641.D
Acq On : 23 Mar 2022 02:56
Operator : JC/MD
Sample : N2062-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Mar 23 07:42:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

