

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027644.D
 Acq On : 23 Mar 2022 04:07
 Operator : JC/MD
 Sample : N2062-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-03-21

Quant Time: Mar 23 07:43:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	356350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	583845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	527005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	243325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	207383	47.241	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.480%
35) Dibromofluoromethane	5.391	113	187410	49.891	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.780%
50) Toluene-d8	8.653	98	678732	47.749	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.500%
62) 4-Bromofluorobenzene	11.079	95	243796	46.663	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.320%
Target Compounds						
16) Acetone	2.379	43	4228	2.363	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027644.D
Acq On : 23 Mar 2022 04:07
Operator : JC/MD
Sample : N2062-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-03-21

Quant Time: Mar 23 07:43:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

