

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027648.D
 Acq On : 23 Mar 2022 05:41
 Operator : JC/MD
 Sample : N2028-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP031822

Quant Time: Mar 23 07:44:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	371185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	579386	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	521708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	243573	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	204218	44.661	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.320%
35) Dibromofluoromethane	5.391	113	188013	50.437	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.880%
50) Toluene-d8	8.653	98	673621	47.755	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.500%
62) 4-Bromofluorobenzene	11.079	95	244621	47.181	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.360%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5115	2.744	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	4669	0.401	ug/l #	37
30) Chloroform	5.111	83	9245	1.308	ug/l #	79
31) Cyclohexane	5.471	56	1474	0.260	ug/l #	79
39) Methylcyclohexane	7.385	83	1334	0.208	ug/l	97
40) Benzene	6.044	78	18244	1.248	ug/l	100
52) Toluene	8.720	92	22495	2.297	ug/l	100
67) Ethyl Benzene	10.201	91	11592	0.631	ug/l	99
68) m/p-Xylenes	10.299	106	22856	3.160	ug/l	98
69) o-Xylene	10.646	106	7294	1.039	ug/l	94
80) 1,3,5-Trimethylbenzene	11.457	105	6223	0.417	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	18776	1.261	ug/l	99
95) Naphthalene	13.780	128	4968	0.294	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027648.D
Acq On : 23 Mar 2022 05:41
Operator : JC/MD
Sample : N2028-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP031822

Quant Time: Mar 23 07:44:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

