

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027649.D
 Acq On : 23 Mar 2022 06:05
 Operator : JC/MD
 Sample : N2028-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB031722

Quant Time: Mar 23 07:45:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	352449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	586394	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	522364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	241175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	205138	47.247	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.500%
35) Dibromofluoromethane	5.391	113	183880	48.739	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.480%
50) Toluene-d8	8.653	98	682082	47.776	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.560%
62) 4-Bromofluorobenzene	11.085	95	245313	46.749	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.500%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4049	2.288	ug/l	93
40) Benzene	6.044	78	6605	0.447	ug/l	97
52) Toluene	8.720	92	8752	0.883	ug/l	99
67) Ethyl Benzene	10.195	91	4551	0.247	ug/l	95
68) m/p-Xylenes	10.305	106	7726	1.067	ug/l	96
69) o-Xylene	10.646	106	2240	0.319	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	5527	0.375	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027649.D
Acq On : 23 Mar 2022 06:05
Operator : JC/MD
Sample : N2028-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB031722

Quant Time: Mar 23 07:45:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

