

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027626.D
 Acq On : 22 Mar 2022 19:49
 Operator : JC/MD
 Sample : N2028-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-116D

Quant Time: Mar 23 01:50:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	355535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	577228	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	499470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	217234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	186223	42.518	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.040%
35) Dibromofluoromethane	5.391	113	184195	49.598	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.200%
50) Toluene-d8	8.653	98	649416	46.211	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.420%
62) 4-Bromofluorobenzene	11.085	95	226233	43.798	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.600%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8864	4.965	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	7462	0.670	ug/l #	84
40) Benzene	6.044	78	7355	0.505	ug/l	95
52) Toluene	8.720	92	9175	0.940	ug/l	96
67) Ethyl Benzene	10.195	91	12821	0.729	ug/l	96
68) m/p-Xylenes	10.305	106	11701	1.690	ug/l	96
69) o-Xylene	10.647	106	2992	0.445	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	3682	0.276	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	10976	0.826	ug/l	96
95) Naphthalene	13.780	128	3091	0.205	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
Data File : VX027626.D
Acq On : 22 Mar 2022 19:49
Operator : JC/MD
Sample : N2028-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-116D

Quant Time: Mar 23 01:50:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49




Instrument :
MSVOA_X
ClientSampleId :
MW-116D


Tgt Ion:168 Resp: 355535
Ion Ratio Lower Upper
168 100
99 53.5 41.7 62.5



#16
Acetone
Concen: 4.965 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.006 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion: 43 Resp: 8864
Ion Ratio Lower Upper
43 100
58 31.7 28.8 43.2









#19
Methyl tert-butyl Ether
Concen: 0.670 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49




Instrument :
MSVOA_X
ClientSampleId :
MW-116D


Tgt Ion: 73 Resp: 7462
Ion Ratio Lower Upper
73 100
57 28.7 16.8 25.2#



#33
1,2-Dichloroethane-d4
Concen: 42.518 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion: 65 Resp: 186223
Ion Ratio Lower Upper
65 100
67 53.8 0.0 107.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49




Instrument :
MSVOA_X
ClientSampleId :
MW-116D


Tgt Ion:114 Resp: 577228
Ion Ratio Lower Upper
114 100
63 18.1 0.0 36.8
88 15.7 0.0 32.4



#35
Dibromofluoromethane
Concen: 49.598 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion:113 Resp: 184195
Ion Ratio Lower Upper
113 100
111 102.2 82.8 124.2
192 20.9 16.9 25.3









#40
Benzene
Concen: 0.505 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49




Instrument :
MSVOA_X
ClientSampleId :
MW-116D


Tgt Ion: 78 Resp: 7355
Ion Ratio Lower Upper
78 100
77 21.7 19.3 28.9



#50
Toluene-d8
Concen: 46.211 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion: 98 Resp: 649416
Ion Ratio Lower Upper
98 100
100 65.3 52.6 78.8





#52

Toluene

Concen: 0.940 ug/l

RT: 8.720 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX027626.D

Acq: 22 Mar 2022 19:49

Instrument :

MSVOA_X

ClientSampleId :





MW-116D

Tgt Ion: 92 Resp: 9175

Ion Ratio Lower Upper

92 100

91 176.9 137.5 206.3



#62

4-Bromofluorobenzene

Concen: 43.798 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.000 min

Lab File: VX027626.D

Acq: 22 Mar 2022 19:49




Tgt Ion: 95 Resp: 226233


Ion Ratio Lower Upper

95 100

174 83.6 0.0 163.0

176 80.0 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49




Instrument :
MSVOA_X
ClientSampleId :
MW-116D


Tgt Ion:117 Resp: 499470
Ion Ratio Lower Upper
117 100
82 54.4 41.8 62.8
119 32.2 24.8 37.2



#67
Ethyl Benzene
Concen: 0.729 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion: 91 Resp: 12821
Ion Ratio Lower Upper
91 100
106 32.8 24.6 36.8









#68
m/p-Xylenes
Concen: 1.690 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49




Instrument :
MSVOA_X
ClientSampleId :
MW-116D


Tgt Ion:106 Resp: 11701
Ion Ratio Lower Upper
106 100
91 207.2 160.4 240.6



#69
o-Xylene
Concen: 0.445 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion:106 Resp: 2992
Ion Ratio Lower Upper
106 100
91 208.5 106.1 318.1









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.024 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX027626.D
 Acq: 22 Mar 2022 19:49




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-116D


Tgt Ion:152 Resp: 217234
 Ion Ratio Lower Upper
 152 100
 115 60.0 37.5 112.7
 150 157.7 0.0 338.2



#80
 1,3,5-Trimethylbenzene
 Concen: 0.276 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. 0.000 min
 Lab File: VX027626.D
 Acq: 22 Mar 2022 19:49

Tgt Ion:105 Resp: 3682
 Ion Ratio Lower Upper
 105 100
 120 53.1 24.3 72.9









#84
1,2,4-Trimethylbenzene
Concen: 0.826 ug/l
RT: 11.756 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Instrument :
MSVOA_X
ClientSampleId :
MW-116D

Tgt Ion:105 Resp: 10976
Ion Ratio Lower Upper
105 100
120 42.8 22.6 67.8



#95
Naphthalene
Concen: 0.205 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX027626.D
Acq: 22 Mar 2022 19:49

Tgt Ion:128 Resp: 3091
Ion Ratio Lower Upper
128 100
127 12.1 10.2 15.4
129 11.3 8.7 13.1

