



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032222\
 Data File : VX027606.D
 Acq On : 22 Mar 2022 09:49
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Title : SW846 8260
 Last Update : Thu Mar 17 15:02:59 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	12441	PASS
75	95	30	60	47.0	36635	PASS
95	95	100	100	100.0	77971	PASS
96	95	5	9	6.8	5326	PASS
173	174	0.00	2	0.9	562	PASS
174	95	50	100	79.4	61872	PASS
175	174	5	9	7.3	4527	PASS
176	174	95	101	95.7	59240	PASS
177	176	5	9	6.7	3976	PASS