

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
 Data File : VX034664.D
 Acq On : 22 Mar 2023 11:26
 Operator : JC/MD
 Sample : VX0322WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0322WBL02

Quant Time: Mar 23 07:52:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	248960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	430178	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	380148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	178149	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	184629	50.976	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.960%
35) Dibromofluoromethane	5.379	113	136360	48.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.020%
50) Toluene-d8	8.647	98	496591	46.836	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.680%
62) 4-Bromofluorobenzene	11.079	95	188903	44.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
Data File : VX034664.D
Acq On : 22 Mar 2023 11:26
Operator : JC/MD
Sample : VX0322WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0322WBL02

Quant Time: Mar 23 07:52:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

