

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
 Data File : VX034670.D
 Acq On : 22 Mar 2023 14:25
 Operator : JC/MD
 Sample : 01941-09 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20230314

Quant Time: Mar 23 07:53:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	303578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	508157	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	446994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	211765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	210930	47.760	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.520%
35) Dibromofluoromethane	5.379	113	154007	45.902	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.800%
50) Toluene-d8	8.647	98	590342	47.134	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.260%
62) 4-Bromofluorobenzene	11.079	95	212050	41.915	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.820%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	6056	1.795	ug/l	93
12) 1,1-Dichloroethene	2.313	96	2617	0.787	ug/l #	67
27) cis-1,2-Dichloroethene	4.495	96	1365	0.321	ug/l	85
44) Trichloroethene	7.123	130	439388	112.178	ug/l	98
64) Tetrachloroethene	9.269	164	1298	0.432	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
Data File : VX034670.D
Acq On : 22 Mar 2023 14:25
Operator : JC/MD
Sample : 01941-09 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20230314

Quant Time: Mar 23 07:53:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

