

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032223\  
 Data File : VX034673.D  
 Acq On : 22 Mar 2023 15:36  
 Operator : JC/MD  
 Sample : 02034-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TB-01-032123

Quant Time: Mar 23 07:54:33 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031023W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Mar 13 10:29:32 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 324904   | 50.000   | ug/l  | 0.01     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 557481   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 485219   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 221053   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 250807   | 53.062   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 106.120% |
| 35) Dibromofluoromethane    | 5.391  | 113   | 170536   | 46.332   | ug/l  | 0.01     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 92.660%  |
| 50) Toluene-d8              | 8.647  | 98    | 649205   | 47.248   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 94.500%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 250987   | 45.222   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 90.440%  |
| Target Compounds            |        |       |          |          |       |          |
|                             |        |       |          |          |       | Qvalue   |
| 44) Trichloroethene         | 7.135  | 130   | 1442     | 0.336    | ug/l  | 100      |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032223\  
Data File : VX034673.D  
Acq On : 22 Mar 2023 15:36  
Operator : JC/MD  
Sample : 02034-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TB-01-032123

Quant Time: Mar 23 07:54:33 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031023W.M  
Quant Title : SW846 8260  
QLast Update : Mon Mar 13 10:29:32 2023  
Response via : Initial Calibration

