

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
 Data File : VX034676.D
 Acq On : 22 Mar 2023 16:46
 Operator : JC/MD
 Sample : 02034-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-CB05-30-35-032123

Quant Time: Mar 23 07:55:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	321534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	546077	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	483575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	236650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	235240	50.290	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.580%	
35) Dibromofluoromethane	5.385	113	162522	45.077	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.160%	
50) Toluene-d8	8.647	98	617195	45.856	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 91.720%#	
62) 4-Bromofluorobenzene	11.079	95	243290	44.751	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.500%	
Target Compounds						Qvalue
16) Acetone	2.380	43	12016	5.453	ug/l	# 88
18) Methyl Acetate	2.703	43	4704	0.653	ug/l	91
20) Methylene Chloride	2.788	84	1262	0.295	ug/l	90
24) 1,1-Dichloroethane	3.611	63	9619	1.315	ug/l	96
25) 2-Butanone	4.574	43	27971	8.872	ug/l	92
30) Chloroform	5.093	83	98695	13.766	ug/l	96
47) Bromodichloromethane	7.824	83	18650	3.351	ug/l	96
60) Dibromochloromethane	9.525	129	5182	1.241	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
Data File : VX034676.D
Acq On : 22 Mar 2023 16:46
Operator : JC/MD
Sample : 02034-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-CB05-30-35-032123

Quant Time: Mar 23 07:55:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

