

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032223\  
 Data File : VX034677.D  
 Acq On : 22 Mar 2023 17:10  
 Operator : JC/MD  
 Sample : 02039-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-7

Quant Time: Mar 23 07:55:27 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031023W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Mar 13 10:29:32 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 325872   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 561735   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 491814   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 225990   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 252534   | 53.268   | ug/l  | 0.01     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 106.540% |
| 35) Dibromofluoromethane    | 5.385  | 113   | 176064   | 47.471   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 94.940%  |
| 50) Toluene-d8              | 8.647  | 98    | 659524   | 47.635   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 95.280%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 252905   | 45.223   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 90.440%  |
| Target Compounds            |        |       |          |          |       |          |
|                             |        |       |          |          |       | Qvalue   |
| 30) Chloroform              | 5.092  | 83    | 11431    | 1.573    | ug/l  | # 82     |
| 64) Tetrachloroethene       | 9.268  | 164   | 1281     | 0.388    | ug/l  | 88       |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032223\  
Data File : VX034677.D  
Acq On : 22 Mar 2023 17:10  
Operator : JC/MD  
Sample : 02039-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-7

Quant Time: Mar 23 07:55:27 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031023W.M  
Quant Title : SW846 8260  
QLast Update : Mon Mar 13 10:29:32 2023  
Response via : Initial Calibration

