

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
 Data File : VX034679.D
 Acq On : 22 Mar 2023 17:57
 Operator : JC/MD
 Sample : 02021-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-46B-032023

Quant Time: Mar 23 07:55:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	331078	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	561422	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	503770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	228813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	252767	52.479	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.960%
35) Dibromofluoromethane	5.385	113	176227	47.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.080%
50) Toluene-d8	8.646	98	666470	48.164	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.320%
62) 4-Bromofluorobenzene	11.079	95	257675	46.101	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.200%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	69514	30.639	ug/l	94
20) Methylene Chloride	2.788	84	31751	7.211	ug/l	97
37) Ethyl Acetate	4.714	43	67128	10.793	ug/l	98
43) Isopropyl Acetate	6.342	43	279649	26.540	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
Data File : VX034679.D
Acq On : 22 Mar 2023 17:57
Operator : JC/MD
Sample : 02021-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-46B-032023

Quant Time: Mar 23 07:55:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

