

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
 Data File : VX034663.D
 Acq On : 22 Mar 2023 10:12
 Operator : JC/MD
 Sample : VX0322WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0322WBL01

Quant Time: Mar 23 07:52:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	251006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	484250	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	429729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169439	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	169006	46.282	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.560%
35) Dibromofluoromethane	5.385	113	132237	41.360	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.720%
50) Toluene-d8	8.647	98	581852	48.750	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.079	95	181071	37.558	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.120%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032223\
Data File : VX034663.D
Acq On : 22 Mar 2023 10:12
Operator : JC/MD
Sample : VX0322WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0322WBL01

Quant Time: Mar 23 07:52:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12




Instrument :
MSVOA_X
ClientSampleId :
VX0322WBL01


Tgt Ion:168 Resp: 251006
Ion Ratio Lower Upper
168 100
99 60.3 45.7 68.5



#33
1,2-Dichloroethane-d4
Concen: 46.282 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.006 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12

Tgt Ion: 65 Resp: 169006
Ion Ratio Lower Upper
65 100
67 51.5 0.0 104.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12




Instrument :
MSVOA_X
ClientSampleId :
VX0322WBL01


Tgt Ion:114 Resp: 484250
Ion Ratio Lower Upper
114 100
63 22.9 0.0 42.2
88 17.3 0.0 33.2



#35
Dibromofluoromethane
Concen: 41.360 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12

Tgt Ion:113 Resp: 132237
Ion Ratio Lower Upper
113 100
111 101.6 83.7 125.5
192 19.3 15.4 23.2









#50
Toluene-d8
Concen: 48.750 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12




Instrument :
MSVOA_X
ClientSampleId :
VX0322WBL01


Tgt Ion: 98 Resp: 581852
Ion Ratio Lower Upper
98 100
100 65.9 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 37.558 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12

Tgt Ion: 95 Resp: 181071
Ion Ratio Lower Upper
95 100
174 75.8 0.0 147.4
176 73.9 0.0 140.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12

Instrument :
MSVOA_X
ClientSampleId :
VX0322WBL01

Tgt Ion:117 Resp: 429729
Ion Ratio Lower Upper
117 100
82 58.1 45.0 67.4
119 32.6 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX034663.D
Acq: 22 Mar 2023 10:12

Tgt Ion:152 Resp: 169439
Ion Ratio Lower Upper
152 100
115 60.0 45.0 135.0
150 158.2 0.0 352.4

