

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\
 Data File : VX008355.D
 Acq On : 23 Mar 2019 17:23
 Operator : JC/SP
 Sample : VX0323WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0323WBL01

Quant Time: Mar 25 08:38:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	230367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	393315	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	331341	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129679	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	165789	47.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.38%	
35) Dibromofluoromethane	5.49	113	122134	47.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.86%	
50) Toluene-d8	8.71	98	493381	48.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.90%	
62) 4-Bromofluorobenzene	11.13	95	160463	43.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008355.D
Acq On : 23 Mar 2019 17:23
Operator : JC/SP
Sample : VX0323WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0323WBL01

Quant Time: Mar 25 08:38:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

