

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\
 Data File : VX008361.D
 Acq On : 23 Mar 2019 19:43
 Operator : JC/SP
 Sample : K1937-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW314S-20190313

Quant Time: Mar 25 08:57:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	218270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	381349	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	319116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126960	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	163337	49.06	ug/l	0.00
Spiked Amount			Recovery	=	98.12%	
35) Dibromofluoromethane	5.50	113	119351	48.31	ug/l	0.00
Spiked Amount			Recovery	=	96.62%	
50) Toluene-d8	8.71	98	474979	48.60	ug/l	0.00
Spiked Amount			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.13	95	156249	43.55	ug/l	0.00
Spiked Amount			Recovery	=	87.10%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008361.D
Acq On : 23 Mar 2019 19:43
Operator : JC/SP
Sample : K1937-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW314S-20190313

Quant Time: Mar 25 08:57:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

