

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\
 Data File : VX008363.D
 Acq On : 23 Mar 2019 20:29
 Operator : JC/SP
 Sample : K1937-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MH-SW4001-SOUTH-20190314

Quant Time: Mar 25 09:01:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	227052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	396847	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	337973	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138005	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	170656	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
35) Dibromofluoromethane	5.50	113	123141	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	8.71	98	497921	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	11.13	95	165978	44.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.90%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008363.D
Acq On : 23 Mar 2019 20:29
Operator : JC/SP
Sample : K1937-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MH-SW4001-SOUTH-20190314

Quant Time: Mar 25 09:01:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

