

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\
 Data File : VX008365.D
 Acq On : 23 Mar 2019 21:16
 Operator : JC/SP
 Sample : K1937-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW307S-20190314

Quant Time: Mar 25 09:04:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221394	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387902	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	330684	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135011	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	166855	49.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.82%	
35) Dibromofluoromethane	5.50	113	120205	47.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.66%	
50) Toluene-d8	8.71	98	485955	48.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.78%	
62) 4-Bromofluorobenzene	11.13	95	161621	44.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008365.D
Acq On : 23 Mar 2019 21:16
Operator : JC/SP
Sample : K1937-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307S-20190314

Quant Time: Mar 25 09:04:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

