

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\
 Data File : VX008366.D
 Acq On : 23 Mar 2019 21:39
 Operator : JC/SP
 Sample : K1937-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW306S-20190314

Quant Time: Mar 25 09:07:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	222183	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	383314	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	325890	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134650	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	165624	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
35) Dibromofluoromethane	5.50	113	120201	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
50) Toluene-d8	8.71	98	480798	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
62) 4-Bromofluorobenzene	11.13	95	160299	44.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.90%	
Target Compounds						
44) Trichloroethene	7.21	130	1922	0.694	ug/l	Qvalue 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008366.D
Acq On : 23 Mar 2019 21:39
Operator : JC/SP
Sample : K1937-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW306S-20190314

Quant Time: Mar 25 09:07:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

