

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\  
 Data File : VX008368.D  
 Acq On : 23 Mar 2019 22:26  
 Operator : JC/SP  
 Sample : K1937-18  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BPS1-TT-MW202S-20190314

Quant Time: Mar 25 09:10:48 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X032319W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Mar 23 15:27:34 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.66  | 168  | 218678   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.86  | 114  | 378790   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.11 | 117  | 325791   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.08 | 152  | 133791   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds |        |     |          |       |        |      |
|-----------------------------|--------|-----|----------|-------|--------|------|
| 33) 1,2-Dichloroethane-d4   | 6.06   | 65  | 163254   | 48.95 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 97.90% |      |
| 35) Dibromofluoromethane    | 5.50   | 113 | 118596   | 48.32 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 96.64% |      |
| 50) Toluene-d8              | 8.71   | 98  | 478061   | 49.25 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 98.50% |      |
| 62) 4-Bromofluorobenzene    | 11.13  | 95  | 158915   | 44.59 | ug/l   | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 89.18% |      |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\  
Data File : VX008368.D  
Acq On : 23 Mar 2019 22:26  
Operator : JC/SP  
Sample : K1937-18  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BPS1-TT-MW202S-20190314

Quant Time: Mar 25 09:10:48 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X032319W.M  
Quant Title : SW846 8260  
QLast Update : Sat Mar 23 15:27:34 2019  
Response via : Initial Calibration

