

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032319\
 Data File : VX008361.D
 Acq On : 23 Mar 2019 19:43
 Operator : JC/SP
 Sample : K1937-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW314S-20190313

Quant Time: Mar 25 08:57:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	218270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	381349	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	319116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126960	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	163337	49.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.12%	
35) Dibromofluoromethane	5.50	113	119351	48.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.62%	
50) Toluene-d8	8.71	98	474979	48.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.13	95	156249	43.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.10%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008361.D
Acq On : 23 Mar 2019 19:43
Operator : JC/SP
Sample : K1937-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW314S-20190313

Quant Time: Mar 25 08:57:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX008361.D

Acq: 23 Mar 2019 19:43

Instrument :

MSVOA_X

ClientSampleId :


BPS1-TT-MW314S-20190313

Tot Ion:168 Resp: 218270

Ion Ratio Lower Upper




168 100

99 61.1 37.6 56.4#



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 49.064 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX008361.D


Acq: 23 Mar 2019 19:43

Tgt Ion: 65 Resp: 163337

Ion Ratio Lower Upper



65 100


67 53.6 0.0 110.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX008361.D
 Acq: 23 Mar 2019 19:43




Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW314S-20190313


Tot Ion	Ratio	Lower	Upper
114	100		
63	23.5	0.0	39.8
88	16.5	0.0	31.4



#35
 Dibromofluoromethane
 Concen: 48.306 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX008361.D
 Acq: 23 Mar 2019 19:43

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.2	81.4	122.0
192	18.9	19.4	29.0#





#50

Toluene-d8

Concen: 48.604 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX008361.D

Acq: 23 Mar 2019 19:43

Instrument :

MSVOA_X

ClientSampleId :


BPS1-TT-MW314S-20190313

Tot Ion: 98 Resp: 474979

Ion Ratio Lower Upper

98 100

100 63.7 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.71



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 43.547 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX008361.D

Acq: 23 Mar 2019 19:43


Tgt Ion: 95 Resp: 156249

Ion Ratio Lower Upper

95 100

174 72.4 0.0 182.8

176 69.7 0.0 178.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX008361.D
Acq: 23 Mar 2019 19:43

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW314S-20190313

Tot Ion:117 Resp: 319116
Ion Ratio Lower Upper
117 100
82 62.6 44.3 66.5
119 31.4 25.3 37.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX008361.D
Acq: 23 Mar 2019 19:43

Tgt Ion:152 Resp: 126960
Ion Ratio Lower Upper
152 100
115 61.9 38.6 115.7
150 157.8 0.0 348.4

