

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX032319\
 Data File : VX008374.D
 Acq On : 24 Mar 2019 00:45
 Operator : JC/SP
 Sample : K1900-08MS
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 11:08:47 AM

Quant Time: Mar 25 09:24:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	201286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	352536	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149458	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	150202	48.93	ug/l	0.00
Spiked Amount			Recovery	=	97.86%	
35) Dibromofluoromethane	5.50	113	111553	48.84	ug/l	0.00
Spiked Amount			Recovery	=	97.68%	
50) Toluene-d8	8.71	98	445272	49.29	ug/l	0.00
Spiked Amount			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	11.13	95	165322	49.84	ug/l	0.00
Spiked Amount			Recovery	=	99.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	107793	47.306	ug/l	100
3) Chloromethane	1.32	50	163688	48.459	ug/l	99
4) Vinyl Chloride	1.40	62	159103	48.259	ug/l	99
5) Bromomethane	1.64	94	100893	48.598	ug/l	97
6) Chloroethane	1.71	64	94149	48.760	ug/l	99
7) Trichlorofluoromethane	1.92	101	176980	47.204	ug/l	100
8) Diethyl Ether	2.18	74	84063	46.827	ug/l	89
9) 1,1,2-Trichlorotrifluoroet	2.38	101	102353	46.085	ug/l	88
10) Methyl Iodide	2.51	142	137783	48.898	ug/l	98
11) Tert butyl alcohol	3.05	59	177807	260.022	ug/l	99
12) 1,1-Dichloroethene	2.37	96	107597	46.932	ug/l	88
13) Acrolein	2.29	56	126378	216.276	ug/l	99
14) Allyl chloride	2.73	41	236893	46.708	ug/l	95
15) Acrylonitrile	3.14	53	428399	244.640	ug/l	99
16) Acetone	2.45	43	369998	241.117	ug/l	98
17) Carbon Disulfide	2.57	76	331050	45.158	ug/l	99
18) Methyl Acetate	2.78	43	155369	40.450	ug/l	98
19) Methyl tert-butyl Ether	3.20	73	402800	48.782	ug/l	98
20) Methylene Chloride	2.85	84	130256	47.152	ug/l	92
21) trans-1,2-Dichloroethene	3.17	96	114333	46.057	ug/l	93
22) Diisopropyl ether	3.85	45	479293	48.536	ug/l #	88
23) Vinyl Acetate	3.81	43	1815667	208.247	ug/l	97
24) 1,1-Dichloroethane	3.70	63	242465	48.158	ug/l	98
25) 2-Butanone	4.67	43	624902	247.314	ug/l	96
26) 2,2-Dichloropropane	4.58	77	123065	37.939	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	133742	47.038	ug/l	92
28) Bromochloromethane	5.01	49	120630	47.995	ug/l #	78
29) Tetrahydrofuran	5.13	42	410452	244.647	ug/l	96
30) Chloroform	5.21	83	219398	48.333	ug/l	99
31) Cyclohexane	5.58	56	229362	47.507	ug/l	95
32) 1,1,1-Trichloroethane	5.49	97	175141	48.768	ug/l	96
36) 1,1-Dichloropropene	5.80	75	171904	47.719	ug/l	95
37) Ethyl Acetate	4.84	43	203754	41.792	ug/l	97
38) Carbon Tetrachloride	5.79	117	145321	47.060	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX032319\
 Data File : VX008374.D
 Acq On : 24 Mar 2019 00:45
 Operator : JC/SP
 Sample : K1900-08MS
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 BPS1-TT-MW309D-20190313MS

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 11:08:47 AM

Quant Time: Mar 25 09:24:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	202349	45.826	ug/l	93
40) Benzene	6.14	78	519453	47.349	ug/l	98
41) Methacrylonitrile	5.04	41	132351	49.804	ug/l	96
42) 1,2-Dichloroethane	6.19	62	188201	47.271	ug/l	96
43) Isopropyl Acetate	6.45	43	337305	44.429	ug/l	97
44) Trichloroethene	7.21	130	117787	46.233	ug/l	88
45) 1,2-Dichloropropane	7.51	63	148563	47.834	ug/l	98
46) Dibromomethane	7.66	93	85945	46.629	ug/l #	84
47) Bromodichloromethane	7.90	83	168438	47.911	ug/l	100
48) Methyl methacrylate	7.77	41	188308	49.155	ug/l	94
49) 1,4-Dioxane	7.74	88	77032	996.639	ug/l	92
51) 4-Methyl-2-Pentanone	8.64	43	1241662	249.271	ug/l	97
52) Toluene	8.78	92	312697	48.608	ug/l	98
53) t-1,3-Dichloropropene	9.04	75	182457	48.270	ug/l	99
54) cis-1,3-Dichloropropene	8.43	75	202118	47.162	ug/l #	91
55) 1,1,2-Trichloroethane	9.21	97	126876	47.482	ug/l	97
56) Ethyl methacrylate	9.18	69	237468	50.199	ug/l	91
57) 1,3-Dichloropropane	9.37	76	226794	47.643	ug/l	100
59) 2-Hexanone	9.49	43	978658	255.936	ug/l	96
60) Dibromochloromethane	9.58	129	121194	48.863	ug/l	98
61) 1,2-Dibromoethane	9.67	107	130748	48.213	ug/l	99
64) Tetrachloroethene	9.34	164	94571	45.796	ug/l	96
65) Chlorobenzene	10.14	112	315219	47.597	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.22	131	111172	48.406	ug/l	98
67) Ethyl Benzene	10.25	91	598972	48.301	ug/l	97
68) m/p-Xylenes	10.36	106	432309	97.027	ug/l	92
69) o-Xylene	10.69	106	212165	48.697	ug/l	93
70) Styrene	10.71	104	372576	49.310	ug/l	96
71) Bromoform	10.85	173	87833	49.646	ug/l #	99
73) Isopropylbenzene	11.02	105	556706	47.199	ug/l	98
74) N-amyl acetate	10.90	43	299092	43.311	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.27	83	222536	47.367	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	210906m	53.912	ug/l	
77) Bromobenzene	11.26	156	131550	47.136	ug/l	80
78) n-propylbenzene	11.36	91	658931	47.665	ug/l	95
79) 2-Chlorotoluene	11.42	91	393411	46.718	ug/l	94
80) 1,3,5-Trimethylbenzene	11.51	105	474123	47.253	ug/l	97
81) trans-1,4-Dichloro-2-buten	11.07	75	64076	46.676	ug/l	91
82) 4-Chlorotoluene	11.51	91	456426	47.040	ug/l	94
83) tert-Butylbenzene	11.77	119	464446	48.609	ug/l	94
84) 1,2,4-Trimethylbenzene	11.80	105	482751	48.212	ug/l	97
85) sec-Butylbenzene	11.94	105	543866	47.494	ug/l	96
86) p-Isopropyltoluene	12.06	119	480743	47.629	ug/l	96
87) 1,3-Dichlorobenzene	12.02	146	234761	46.835	ug/l	97
88) 1,4-Dichlorobenzene	12.10	146	236457	46.363	ug/l	97
89) n-Butylbenzene	12.38	91	454330	47.150	ug/l	97
90) Hexachloroethane	12.60	117	80090	48.037	ug/l	80
91) 1,2-Dichlorobenzene	12.39	146	239944	47.265	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	13.00	75	49736	49.670	ug/l	76
93) 1,2,4-Trichlorobenzene	13.65	180	160959	47.019	ug/l	98

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032319\
Data File : VX008374.D
Acq On : 24 Mar 2019 00:45
Operator : JC/SP
Sample : K1900-08MS
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS


Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM

Quant Time: Mar 25 09:24:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
94) Hexachlorobutadiene	13.78	225	73772	45.600	ug/l	99
95) Naphthalene	13.83	128	575305	47.856	ug/l	100
96) 1,2,3-Trichlorobenzene	14.02	180	161732	46.819	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:168 Resp: 201286

Ion Ratio Lower Upper

168 100





99 57.4 37.6 56.4#

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#2

Dichlorodifluoromethane

Concen: 47.306 ug/l

RT: 1.20 min Scan# 7

Delta R.T. -0.00 min

Lab File: VX008374.D




Acq: 24 Mar 2019 00:45


Tgt Ion: 85 Resp: 107793

Ion Ratio Lower Upper

85 100

87 33.0 16.4 49.1









#3
Chloromethane
Concen: 48.459 ug/l
RT: 1.32 min Scan# 28
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion: 50 Resp: 163688
Ion Ratio Lower Upper
50 100
52 32.3 26.5 39.7




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#4
Vinyl Chloride
Concen: 48.259 ug/l
RT: 1.40 min Scan# 41
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion: 62 Resp: 159103
Ion Ratio Lower Upper
62 100
64 31.9 25.3 37.9





#5

Bromomethane

Concen: 48.598 ug/l

RT: 1.64 min Scan# 79

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 94 Resp: 100893

Ion Ratio Lower Upper

94 100





96 91.8 75.5 113.3

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#6

Chloroethane

Concen: 48.760 ug/l

RT: 1.71 min Scan# 92

Delta R.T. 0.00 min

Lab File: VX008374.D




Acq: 24 Mar 2019 00:45


Tgt Ion: 64 Resp: 94149

Ion Ratio Lower Upper

64 100

66 31.9 25.2 37.8





#7





Trichlorofluoromethane
Concen: 47.204 ug/l
RT: 1.92 min Scan# 126
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
101	100		
103	62.9	50.4	75.6

Manual Integrations
APPROVED




MMDadoda
3/25/2019 11:08:47 AM




#8

Diethyl Ether
Concen: 46.827 ug/l
RT: 2.18 min Scan# 169
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
74	100		
45	115.5	52.1	156.4





#9

1,1,2-Trichlorotrifluoroethane

Concen: 46.085 ug/l

RT: 2.38 min Scan# 201

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:	101	Resp:	102353
Ion Ratio	Lower	Upper	
101	100		
85	45.0	33.8	50.6
151	72.6	69.1	103.7

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#10

Methyl Iodide

Concen: 48.898 ug/l




RT: 2.51 min Scan# 222


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion:	142	Resp:	137783
Ion Ratio	Lower	Upper	
142	100		
127	42.0	32.1	48.1
141	14.2	11.4	17.2





#11





Tert butyl alcohol
Concen: 260.022 ug/l
RT: 3.05 min Scan# 311
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion: 59 Resp: 177807
Ion Ratio Lower Upper
59 100
57 9.8 8.2 12.2

Manual Integrations
APPROVED




MMDadoda
3/25/2019 11:08:47 AM




#12

1,1-Dichloroethene
Concen: 46.932 ug/l
RT: 2.37 min Scan# 199
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion: 96 Resp: 107597
Ion Ratio Lower Upper
96 100
61 184.3 129.8 194.6
98 64.0 51.0 76.6





#13

Acrolein

Concen: 216.276 ug/l

RT: 2.29 min Scan# 187

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 56 Resp: 126378

Ion Ratio Lower Upper

56 100





55 71.4 57.8 86.6

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#14

Allyl chloride

Concen: 46.708 ug/l

RT: 2.73 min Scan# 258

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45




Tgt Ion: 41 Resp: 236893


Ion Ratio Lower Upper

41 100

39 61.9 46.1 69.1

76 27.7 24.0 36.0





#15

Acrylonitrile

Concen: 244.640 ug/l

RT: 3.14 min Scan# 326

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:	53	Resp:	428399
Ion	Ratio	Lower	Upper
53	100		
52	83.0	66.0	99.0
51	35.8	28.0	42.0

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#16

Acetone

Concen: 241.117 ug/l




RT: 2.45 min Scan# 212


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion:	43	Resp:	369998
Ion	Ratio	Lower	Upper
43	100		
58	30.8	25.5	38.3





#17

Carbon Disulfide

Concen: 45.158 ug/l

RT: 2.57 min Scan# 232

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 76 Resp: 331050

Ion Ratio Lower Upper

76 100


78 9.3 7.2 10.8

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.57

200000



150000

100000

50000

0

Time-->



#18

Methyl Acetate

Concen: 40.450 ug/l

RT: 2.78 min Scan# 266

Delta R.T. -0.00 min

Lab File: VX008374.D


Acq: 24 Mar 2019 00:45

Tgt Ion: 43 Resp: 155369

Ion Ratio Lower Upper

43 100

74 21.2 17.8 26.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0

2.78

100000

80000


60000


40000

20000

0

Time-->









#19
Methyl tert-butyl Ether
Concen: 48.782 ug/l
RT: 3.20 min Scan# 335
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion: 73 Resp: 402800
Ion Ratio Lower Upper
73 100
57 23.7 18.2 27.4




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#20
Methylene Chloride
Concen: 47.152 ug/l
RT: 2.85 min Scan# 279
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion: 84 Resp: 130256
Ion Ratio Lower Upper
84 100
49 139.0 99.3 148.9
51 42.6 31.3 46.9
86 64.9 52.6 79.0









#21
trans-1,2-Dichloroethene
Concen: 46.057 ug/l
RT: 3.17 min Scan# 330
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
96	100		
61	154.9	114.4	171.6
98	64.7	51.1	76.7




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#22
Diisopropyl ether
Concen: 48.536 ug/l
RT: 3.85 min Scan# 443
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
45	100		
43	68.5	45.3	67.9#
87	23.2	21.4	32.0
59	10.9	9.5	14.3





#23

Vinyl Acetate

Concen: 208.247 ug/l

RT: 3.81 min Scan# 436

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 43 Resp: 1815667

Ion Ratio Lower Upper

43 100


86 8.7 7.9 11.9

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0

800000

600000



400000

200000

0

3.70 3.80 3.90 4.00 4.10

Time-->



#24

1,1-Dichloroethane

Concen: 48.158 ug/l

RT: 3.70 min Scan# 417

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45


Tgt Ion: 63 Resp: 242465

Ion Ratio Lower Upper

63 100

98 5.7 3.3 9.8

100 3.9 2.3 6.8



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0


100000


50000

0

3.60 3.70 3.80

Time-->





#25

2-Butanone

Concen: 247.314 ug/l

RT: 4.67 min Scan# 577

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 43 Resp: 624902

Ion Ratio Lower Upper

43 100





72 23.1 19.9 29.9

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#26

2,2-Dichloropropane

Concen: 37.939 ug/l

RT: 4.58 min Scan# 562

Delta R.T. -0.00 min

Lab File: VX008374.D




Acq: 24 Mar 2019 00:45


Tgt Ion: 77 Resp: 123065

Ion Ratio Lower Upper

77 100

97 23.6 12.2 36.6





#27





cis-1,2-Dichloroethene
 Concen: 47.038 ug/l
 RT: 4.60 min Scan# 565
 Delta R.T. 0.01 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
96	100		
61	157.9	0.0	288.0
98	63.5	0.0	129.0

Manual Integrations
 APPROVED




MMDadoda
 3/25/2019 11:08:47 AM




#28

Bromochloromethane
 Concen: 47.995 ug/l
 RT: 5.01 min Scan# 633
 Delta R.T. -0.01 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.0
130	61.3	65.0	97.4#









#29
Tetrahydrofuran
Concen: 244.647 ug/l
RT: 5.13 min Scan# 652
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
42	100		
72	39.3	33.8	50.6
71	36.7	31.5	47.3




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#30
Chloroform
Concen: 48.333 ug/l
RT: 5.21 min Scan# 665
Delta R.T. -0.01 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
83	100		
85	64.0	51.8	77.6





#31

Cyclohexane

Concen: 47.507 ug/l

RT: 5.58 min Scan# 726

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:	56	Resp:	229362
Ion	Ratio	Lower	Upper
56	100		
69	29.3	24.0	36.0
84	75.9	65.3	97.9

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#32

1,1,1-Trichloroethane

Concen: 48.768 ug/l




RT: 5.49 min Scan# 712


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion:	97	Resp:	175141
Ion	Ratio	Lower	Upper
97	100		
99	64.4	52.2	78.2
61	49.5	34.9	52.3









#33
 1,2-Dichloroethane-d4
 Concen: 48.925 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion: 65 Resp: 150202
 Ion Ratio Lower Upper
 65 100
 67 53.7 0.0 110.0




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion:114 Resp: 352536
 Ion Ratio Lower Upper
 114 100
 63 23.7 0.0 39.8
 88 16.7 0.0 31.4





#35

Dibromofluoromethane

Concen: 48.840 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.01 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:	113	Resp:	111553
Ion Ratio	Lower	Upper	
113	100		
111	102.1	81.4	122.0
192	19.6	19.4	29.0

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#36

1,1-Dichloropropene

Concen: 47.719 ug/l




RT: 5.80 min Scan# 762


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion:	75	Resp:	171904
Ion Ratio	Lower	Upper	
75	100		
110	32.3	18.4	55.0
77	30.3	25.0	37.4





#37

Ethyl Acetate

Concen: 41.792 ug/l

RT: 4.84 min Scan# 604

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45





Instrument :
MSVOA_X

ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
43	100		
61	13.0	11.3	16.9
70	9.6	8.5	12.7

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#38

Carbon Tetrachloride

Concen: 47.060 ug/l




RT: 5.79 min Scan# 760


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
117	100		
119	92.8	76.2	114.2
121	31.6	24.6	36.8









#39
Methylvlcyclohexane
Concen: 45.826 ug/l
RT: 7.46 min Scan# 1035
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Resp	Lower	Upper
83	100			
55	84.9	60.4	90.6	
98	45.4	36.5	54.7	




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#40
Benzene
Concen: 47.349 ug/l
RT: 6.14 min Scan# 818
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Resp	Lower	Upper
78	100			
77	24.5	19.0	28.4	









#41
Methacrylonitrile
Concen: 49.804 ug/l
RT: 5.04 min Scan# 638
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
41	100		
39	53.6	41.6	62.4
67	64.2	55.8	83.6
52	26.9	21.7	32.5




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#42
1,2-Dichloroethane
Concen: 47.271 ug/l
RT: 6.19 min Scan# 826
Delta R.T. -0.01 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
62	100		
98	8.7	0.0	20.4





#43

Isopropyl Acetate

Concen: 44.429 ug/l

RT: 6.45 min Scan# 868

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X


ClientSampleId :

BPS1-TT-MW309D-20190313MS




Tot Ion:	43	Resp:	337305
Ion	Ratio	Lower	Upper
43	100		
61	19.6	16.3	24.5
87	11.5	10.5	15.7

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 61.00 (60.70 to 61.70): VX0
Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 46.233 ug/l


RT: 7.21 min Scan# 994

Delta R.T. -0.00 min



Lab File: VX008374.D


Acq: 24 Mar 2019 00:45

Tgt Ion:	130	Resp:	117787
Ion	Ratio	Lower	Upper
130	100		
95	105.3	0.0	188.0



Abundance Ion 129.80 (129.50 to 130.50): V
Ion 94.90 (94.60 to 95.60): VX0









#45
 1,2-Dichloropropane
 Concen: 47.834 ug/l
 RT: 7.51 min Scan# 1043
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion: 63 Resp: 148563
 Ion Ratio Lower Upper
 63 100
 65 30.0 24.7 37.1




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#46
 Dibromomethane
 Concen: 46.629 ug/l
 RT: 7.66 min Scan# 1067
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion: 93 Resp: 85945
 Ion Ratio Lower Upper
 93 100
 95 83.4 66.8 100.2
 174 91.3 97.3 145.9#





#47

Bromodichloromethane

Concen: 47.911 ug/l

RT: 7.90 min Scan# 1106

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:	83	Resp:	168438
Ion Ratio	Lower	Upper	
83	100		
85	63.7	50.9	76.3
127	8.3	7.4	11.0

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#48

Methyl methacrylate

Concen: 49.155 ug/l




RT: 7.77 min Scan# 1085


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion:	41	Resp:	188308
Ion Ratio	Lower	Upper	
41	100		
69	75.0	65.2	97.8
39	52.8	39.4	59.2









#49
1,4-Dioxane
Concen: 996.639 ug/l
RT: 7.74 min Scan# 1080
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
88	100		
43	35.9	25.1	37.7
58	79.0	57.8	86.6




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#50
Toluene-d8
Concen: 49.288 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
98	100		
100	64.1	51.7	77.5





#51

4-Methyl-2-Pentanone

Concen: 249.271 ug/l

RT: 8.64 min Scan# 1228

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 43 Resp: 1241662

Ion Ratio Lower Upper

43 100

58 37.5

31.3


46.9

Manual Integrations

APPROVED




MMDadoda

3/25/2019 11:08:47 AM



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 48.608 ug/l

RT: 8.78 min Scan# 1251

Delta R.T. -0.01 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion: 92 Resp: 312697


Ion Ratio Lower Upper

92 100

91 174.7



137.4


206.0



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0









#53
 t-1,3-Dichloropropene
 Concen: 48.270 ug/l
 RT: 9.04 min Scan# 1293
 Delta R.T. -0.01 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion: 75 Resp: 182457
 Ion Ratio Lower Upper
 75 100
 77 31.3 25.4 38.0




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#54
 cis-1,3-Dichloropropene
 Concen: 47.162 ug/l
 RT: 8.43 min Scan# 1194
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion: 75 Resp: 202118
 Ion Ratio Lower Upper
 75 100
 77 31.9 25.0 37.6
 39 55.2 36.6 54.8#






#55
 1,1,2-Trichloroethane
 Concen: 47.482 ug/l
 RT: 9.21 min Scan# 1322
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS




Tot Ion	Ion	Ratio	Resp	Lower	Upper
97	100		126876		
83	88.9	69.1	103.7		
85	57.8	44.3	66.5		
99	61.0	49.8	74.8		

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 11:08:47 AM




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 50.199 ug/l
 RT: 9.18 min Scan# 1316
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
69	100		237468		
41	74.3	54.2	81.4		
39	36.9	25.0	37.4		



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0









#57
 1,3-Dichloropropane
 Concen: 47.643 ug/l
 RT: 9.37 min Scan# 1348
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion: 76 Resp: 226794
 Ion Ratio Lower Upper
 76 100
 78 32.0 25.5 38.3




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#59
 2-Hexanone
 Concen: 255.936 ug/l
 RT: 9.49 min Scan# 1368
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion: 43 Resp: 978658
 Ion Ratio Lower Upper
 43 100
 58 52.2 27.6 82.7









#60
Dibromochloromethane
Concen: 48.863 ug/l
RT: 9.58 min Scan# 1382
Delta R.T. -0.01 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion:129 Resp: 121194
Ion Ratio Lower Upper
129 100
127 77.8 38.3 114.8




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#61
1,2-Dibromoethane
Concen: 48.213 ug/l
RT: 9.67 min Scan# 1397
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion:107 Resp: 130748
Ion Ratio Lower Upper
107 100
109 94.1 75.8 113.8









#62
4-Bromofluorobenzene
Concen: 49.841 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
95	100		
174	72.9	0.0	182.8
176	71.1	0.0	178.0




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.01 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
117	100		
82	60.1	44.3	66.5
119	32.0	25.3	37.9





#64

Tetrachloroethene

Concen: 45.796 ug/l

RT: 9.34 min Scan# 1342

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:164 Resp: 94571

Ion Ratio Lower Upper

164 100

166 128.8 103.4 155.0

129 97.6 72.2 108.4


131 92.9 69.4 104.2

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM






Abundance

Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V



#65

Chlorobenzene

Concen: 47.597 ug/l

RT: 10.14 min Scan# 1474

Delta R.T. -0.00 min

Lab File: VX008374.D


Acq: 24 Mar 2019 00:45

Tgt Ion:112 Resp: 315219

Ion Ratio Lower Upper

112 100



114 32.3 26.0 39.0




Abundance

Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V









#66
 1,1,1,2-Tetrachloroethane
 Concen: 48.406 ug/l
 RT: 10.22 min Scan# 1487
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion:131 Resp: 111172
 Ion Ratio Lower Upper
 131 100
 133 95.4 47.8 143.3
 119 68.3 32.5 97.5




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#67
 Ethyl Benzene
 Concen: 48.301 ug/l
 RT: 10.25 min Scan# 1492
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion: 91 Resp: 598972
 Ion Ratio Lower Upper
 91 100
 106 29.9 25.2 37.8





#68

m/p-Xylenes

Concen: 97.027 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:106 Resp: 432309

Ion Ratio Lower Upper

106 100





91 211.8 159.8 239.6

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#69

o-Xylene

Concen: 48.697 ug/l

RT: 10.69 min Scan# 1565

Delta R.T. -0.01 min

Lab File: VX008374.D




Acq: 24 Mar 2019 00:45


Tgt Ion:106 Resp: 212165

Ion Ratio Lower Upper

106 100

91 223.2 106.3 318.9






#70
Stvrene
Concen: 49.310 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS




Tot Ion	Ratio	Lower	Upper
104	100		
78	53.7	38.5	57.7
103	55.7	44.1	66.1

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM




Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V






#71
Bromoform
Concen: 49.646 ug/l
RT: 10.85 min Scan# 1591
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
173	100		
175	49.6	24.3	72.9
254	0.2	0.1	0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.08 min Scan# 1792
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
152	100		
115	85.6	38.6	115.7
150	173.5	0.0	348.4

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 11:08:47 AM






Abundance

Ion 151.90 (151.60 to 152.60): V


Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73
 Isopropylbenzene
 Concen: 47.199 ug/l
 RT: 11.02 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45



Tgt Ion	Ratio	Lower	Upper
105	100		
120	25.6	13.4	40.1




Abundance

Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-amvl acetate

Concen: 43.311 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Resp	Lower	Upper
43	100	299092		
70	37.8		33.5	50.3
55	23.0		19.2	28.8
61	23.1		19.0	28.6

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#75

1,1,2,2-Tetrachloroethane

Concen: 47.367 ug/l




RT: 11.27 min Scan# 1659


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Resp	Lower	Upper
83	100	222536		
131	8.9		5.4	16.2
85	63.9		32.3	96.8









#76
 1,2,3-Trichloropropane
 Concen: 53.912 ug/l m
 RT: 11.29 min Scan# 1663
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion: 75 Resp: 210906
 Ion Ratio Lower Upper
 75 100
 77 37.9 18.9 56.7




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#77
 Bromobenzene
 Concen: 47.136 ug/l
 RT: 11.26 min Scan# 1657
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion: 156 Resp: 131550
 Ion Ratio Lower Upper
 156 100
 77 185.7 72.0 215.9
 158 97.7 48.6 145.9





#78

n-propylbenzene

Concen: 47.665 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 91 Resp: 658931

Ion Ratio Lower Upper

91 100





120 21.6 12.0 36.0

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#79

2-Chlorotoluene

Concen: 46.718 ug/l

RT: 11.42 min Scan# 1684

Delta R.T. -0.00 min

Lab File: VX008374.D




Acq: 24 Mar 2019 00:45


Tgt Ion: 91 Resp: 393411

Ion Ratio Lower Upper

91 100

126 31.5 17.5 52.6





#80

1,3,5-Trimethylbenzene

Concen: 47.253 ug/l

RT: 11.51 min Scan# 1698

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:105 Resp: 474123

Ion Ratio Lower Upper

105 100





120 47.3 24.9 74.6

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#81

trans-1,4-Dichloro-2-butene

Concen: 46.676 ug/l

RT: 11.07 min Scan# 1627

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45




Tgt Ion: 75 Resp: 64076


Ion Ratio Lower Upper

75 100

53 106.1 76.1 114.1

89 43.6 36.3 54.5





#82

4-Chlorotoluene

Concen: 47.040 ug/l

RT: 11.51 min Scan# 1699

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion: 91 Resp: 456426

Ion Ratio Lower Upper

91 100





126 27.8 15.6 46.8

Manual Integrations

APPROVED

MMDadoda

3/25/2019 11:08:47 AM



#83

tert-Butylbenzene

Concen: 48.609 ug/l

RT: 11.77 min Scan# 1741

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45




Tgt Ion: 119 Resp: 464446


Ion Ratio Lower Upper

119 100

91 59.0 27.1 81.3

134 21.4 11.4 34.1









#84
 1,2,4-Trimethylbenzene
 Concen: 48.212 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion:105 Resp: 482751
 Ion Ratio Lower Upper
 105 100
 120 43.2 22.6 67.7




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#85
 sec-Butylbenzene
 Concen: 47.494 ug/l
 RT: 11.94 min Scan# 1770
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion:105 Resp: 543866
 Ion Ratio Lower Upper
 105 100
 134 18.7 10.4 31.2









#86
 p-Isopropyltoluene
 Concen: 47.629 ug/l
 RT: 12.06 min Scan# 1789
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Resp	Lower	Upper
119	100	480743		
134	25.3		13.3	39.9
91	24.7		10.9	32.7




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#87
 1,3-Dichlorobenzene
 Concen: 46.835 ug/l
 RT: 12.02 min Scan# 1783
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	234761		
111	41.8		18.9	56.7
148	64.0		31.7	95.1









#88
 1,4-Dichlorobenzene
 Concen: 46.363 ug/l
 RT: 12.10 min Scan# 1795
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Lower	Upper
146	100		
111	41.4	18.4	55.0
148	63.9	32.2	96.6




Manual Integrations
 APPROVED


MMDadoda
 3/25/2019 11:08:47 AM



#89
 n-Butylbenzene
 Concen: 47.150 ug/l
 RT: 12.38 min Scan# 1842
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Tgt Ion	Ratio	Lower	Upper
91	100		
92	53.0	27.0	80.8
134	23.5	13.7	41.0









#90
Hexachloroethane
Concen: 48.037 ug/l
RT: 12.60 min Scan# 1877
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion:117 Resp: 80090
Ion Ratio Lower Upper
117 100
201 78.0 48.9 146.7




Manual Integrations
APPROVED


MMDadoda
3/25/2019 11:08:47 AM



#91
1,2-Dichlorobenzene
Concen: 47.265 ug/l
RT: 12.39 min Scan# 1843
Delta R.T. -0.00 min
Lab File: VX008374.D
Acq: 24 Mar 2019 00:45

Tgt Ion:146 Resp: 239944
Ion Ratio Lower Upper
146 100
111 43.8 19.6 58.8
148 64.0 31.9 95.5





#92

1,2-Dibromo-3-Chloropropane

Concen: 49.670 ug/l

RT: 13.00 min Scan# 1943

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :
MSVOA_X





ClientSampleId :
BPS1-TT-MW309D-20190313MS

Tot Ion: 75 Resp: 49736

Ion	Ratio	Lower	Upper
75	100		
155	76.8	49.7	149.1
157	99.4	63.8	191.4

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#93

1,2,4-Trichlorobenzene

Concen: 47.019 ug/l

RT: 13.65 min Scan# 2049




Delta R.T. -0.00 min


Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion: 180 Resp: 160959

Ion	Ratio	Lower	Upper
180	100		
182	94.4	47.9	143.7
145	30.8	14.1	42.4





#94

Hexachlorobutadiene

Concen: 45.600 ug/l

RT: 13.78 min Scan# 2071

Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Instrument :

MSVOA_X





ClientSampleId :

BPS1-TT-MW309D-20190313MS

Tot Ion:	225	Resp:	73772
Ion Ratio	Lower	Upper	
225	100		
223	62.9	31.1	93.2
227	64.3	32.0	96.0

Manual Integrations
APPROVED

MMDadoda
3/25/2019 11:08:47 AM



#95

Naphthalene

Concen: 47.856 ug/l




RT: 13.83 min Scan# 2080


Delta R.T. -0.00 min

Lab File: VX008374.D

Acq: 24 Mar 2019 00:45

Tgt Ion:	128	Resp:	575305
Ion Ratio	Lower	Upper	
128	100		
127	13.1	10.3	15.5
129	10.8	8.7	13.1





#96
 1,2,3-Trichlorobenzene
 Concen: 46.819 ug/l
 RT: 14.02 min Scan# 2110
 Delta R.T. -0.00 min
 Lab File: VX008374.D
 Acq: 24 Mar 2019 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309D-20190313MS

Tot Ion	Ratio	Resp	Lower	Upper
180	100			
182	94.8	47.5	142.5	
145	32.9	14.8	44.3	

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 11:08:47 AM

