


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX032319\
 Data File : VX008353.D
 Acq On : 23 Mar 2019 15:54
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Title : SW846 8260
 Last Update : Sat Mar 23 15:27:34 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	12776	PASS
75	95	30	60	52.2	31697	PASS
95	95	100	100	100.0	60752	PASS
96	95	5	9	6.7	4050	PASS
173	174	0.00	2	0.7	319	PASS
174	95	50	100	71.8	43626	PASS
175	174	5	9	7.7	3353	PASS
176	174	95	101	95.8	41813	PASS
177	176	5	9	6.6	2755	PASS