

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021375.D
 Acq On : 23 Mar 2021 11:09
 Operator : JC/MD
 Sample : VX0323MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0323MBL01

Quant Time: Mar 23 15:03:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	75022	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.823	114	118727	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	100365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	38436	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.023	65	56933	46.27	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.54%
35) Dibromofluoromethane	5.458	113	42083	52.09	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.18%
50) Toluene-d8	8.694	98	146154	49.86	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.72%
62) 4-Bromofluorobenzene	11.118	95	47961	42.13	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.26%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021375.D
Acq On : 23 Mar 2021 11:09
Operator : JC/MD
Sample : VX0323MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0323MBL01

Quant Time: Mar 23 15:03:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

