

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021379.D
 Acq On : 23 Mar 2021 12:54
 Operator : JC/MD
 Sample : M1723-03DL 10X
 Misc : 11.36g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RO124027DL

Quant Time: Mar 23 15:18:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	77012	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	123061	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	108497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	45566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	58880	46.61	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.22%		
35) Dibromofluoromethane	5.464	113	42222	50.42	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.84%		
50) Toluene-d8	8.694	98	150035	49.38	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.76%		
62) 4-Bromofluorobenzene	11.118	95	53795	45.59	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.18%		
Target Compounds					Qvalue	
16) Acetone	2.423	43	2458	5.29	ug/l #	82
18) Methyl Acetate	2.746	43	4865	3.82	ug/l #	89
51) 4-Methyl-2-Pentanone	8.618	43	6000	4.38	ug/l #	86
52) Toluene	8.765	92	31984	15.14	ug/l	100
56) Ethyl methacrylate	9.159	69	3710	2.67	ug/l #	66
67) Ethyl Benzene	10.236	91	46125	11.18	ug/l	98
68) m/p-Xylenes	10.341	106	49669	33.70	ug/l	93
69) o-Xylene	10.683	106	12818	8.95	ug/l	91
70) Styrene	10.694	104	5693	2.37	ug/l #	61
73) Isopropylbenzene	11.000	105	2906	0.86	ug/l	96
78) n-propylbenzene	11.342	91	5198	1.31	ug/l	99
80) 1,3,5-Trimethylbenzene	11.489	105	3232	1.13	ug/l	99
84) 1,2,4-Trimethylbenzene	11.789	105	6357	2.18	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021379.D
Acq On : 23 Mar 2021 12:54
Operator : JC/MD
Sample : M1723-03DL 10X
Misc : 11.36g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RO124027DL

Quant Time: Mar 23 15:18:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

