

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021380.D
 Acq On : 23 Mar 2021 13:18
 Operator : JC/MD
 Sample : M1701-02DL 10X
 Misc : 5.56g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5R(13-15FT)DL

Quant Time: Mar 23 15:44:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	79233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	129736	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115761	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	53568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62117	47.80	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.60%
35) Dibromofluoromethane	5.464	113	44610	50.53	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.06%
50) Toluene-d8	8.694	98	157018	49.02	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.04%
62) 4-Bromofluorobenzene	11.118	95	61372	49.34	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.68%
Target Compounds						
73) Isopropylbenzene	11.000	105	8896	2.24	ug/l	99
78) n-propylbenzene	11.341	91	25246	5.43	ug/l	95
85) sec-Butylbenzene	11.930	105	45846	12.14	ug/l	94
89) n-Butylbenzene	12.371	91	43390	13.75	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021380.D
Acq On : 23 Mar 2021 13:18
Operator : JC/MD
Sample : M1701-02DL 10X
Misc : 5.56g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5R(13-15FT)DL

Quant Time: Mar 23 15:44:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

