

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021387.D
 Acq On : 23 Mar 2021 15:54
 Operator : JC/MD
 Sample : M1736-02DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20210317DL

Quant Time: Mar 24 01:57:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	85700	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.829	114	139347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	123192	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52185	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	64915	46.18	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.36%
35) Dibromofluoromethane	5.464	113	46765	49.32	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.64%
50) Toluene-d8	8.694	98	168772	49.06	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.12%
62) 4-Bromofluorobenzene	11.118	95	61659	46.15	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.30%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	663	0.74	ug/l	85
12) 1,1-Dichloroethene	2.352	96	411	0.47	ug/l	97
44) Trichloroethene	7.188	130	48679	47.36	ug/l	95
64) Tetrachloroethene	9.312	164	297	0.32	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021387.D
Acq On : 23 Mar 2021 15:54
Operator : JC/MD
Sample : M1736-02DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20210317DL

Quant Time: Mar 24 01:57:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

