

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021388.D
 Acq On : 23 Mar 2021 16:14
 Operator : JC/MD
 Sample : M1750-02 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1015

Quant Time: Mar 24 01:57:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	83621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	137782	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	119266	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	64315	46.89	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.78%
35) Dibromofluoromethane	5.464	113	46217	49.29	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.58%
50) Toluene-d8	8.694	98	165520	48.66	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.32%
62) 4-Bromofluorobenzene	11.118	95	59996	45.41	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.82%
Target Compounds						
11) Tert butyl alcohol	3.023	59	3957	17.17	ug/l	96
16) Acetone	2.423	43	525112	1040.55	ug/l	89
52) Toluene	8.765	92	35974	15.21	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021388.D
Acq On : 23 Mar 2021 16:14
Operator : JC/MD
Sample : M1750-02 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1015

Quant Time: Mar 24 01:57:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

