

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021392.D
 Acq On : 23 Mar 2021 17:43
 Operator : JC/MD
 Sample : PB135291TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135291TB

Quant Time: Mar 24 01:58:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	74719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	123520	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	108108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	44250	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	59356	48.43	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.86%
35) Dibromofluoromethane	5.458	113	41769	49.69	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.38%
50) Toluene-d8	8.694	98	152908	50.14	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.28%
62) 4-Bromofluorobenzene	11.118	95	52557	44.38	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.76%
Target Compounds						
16) Acetone	2.423	43	10013	22.21	ug/l	88
20) Methylene Chloride	2.834	84	1527	1.65	ug/l	87
43) Isopropyl Acetate	6.411	43	10380	4.39	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021392.D
Acq On : 23 Mar 2021 17:43
Operator : JC/MD
Sample : PB135291TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135291TB

Quant Time: Mar 24 01:58:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

