

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021394.D
 Acq On : 23 Mar 2021 18:30
 Operator : JC/MD
 Sample : PB135263TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135263TB

Quant Time: Mar 24 01:59:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	73632	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	120354	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	59198	49.02	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 98.04%			
35) Dibromofluoromethane	5.464	113	41242	50.36	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.72%			
50) Toluene-d8	8.694	98	153204	51.56	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.12%			
62) 4-Bromofluorobenzene	11.118	95	55661	48.23	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.46%			
Target Compounds						Qvalue
11) Tert butyl alcohol	3.011	59	1651	8.14	ug/l	97
16) Acetone	2.417	43	8880	19.98	ug/l #	88
20) Methylene Chloride	2.835	84	6805	7.45	ug/l #	84
25) 2-Butanone	4.635	43	6290	9.21	ug/l	90
43) Isopropyl Acetate	6.406	43	10111	4.39	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021394.D
Acq On : 23 Mar 2021 18:30
Operator : JC/MD
Sample : PB135263TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135263TB

Quant Time: Mar 24 01:59:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

